ARGUS GROUP

CALIBRATION LOG

Instrument: **Serial Number:**

<u>Calibration materials required:</u>

Span Gas Cylinder: 50% LEL Methane / 250 ppm CO / 25 ppm H_2S / balance Air

Cal accessories: 1.0 lpm Regulator, Tubing, and Calibration Adapter

<u>Calibration procedure</u> <u>(see instrument manual for complete instructions)</u>**:**

1. Zero - instrument should read: 0% LEL / 20.9% O_2 / 0 ppm CO / 0 ppm H_2S
2. Span - instrument should read: 55-65% LEL / 16.5-17.5% O_2 / 240-260 ppm CO / 23-27 ppm H_2S

Date	Name	Reading before adjustment		Reading after adjustment		Comments
		Zero	Span	Zero	Span	